

Ankle small FOV16 GE

Indications	Pain, swelling, trauma					
Diagnostic Task	Detects fractures, hematomas, arthritis, bone cyst					
Scan mode	Helical					
Position/Landmark	Head or feet first-supine-include joint of interest only S100-I100					
Topogram	AP 120kV 10mA Lat 120kV 10mA					
kVp/Reference mass	120kv 200mA					
Rotation time/pitch	1.0/0.938:1					
Detector Configuration	16x0.625					
Table Speed/Increment	9.37					
Dose reduction	Noise Index na					
Allowed CTDI ranges*	7mGy-50mGy					
XR29 Dose Notification value	50mGy					
Helical Set		body	thickness			recon
	recon	part	spacing	kernel	window	destination
	1	ankle bone	.625mmx .625mm	bone		pacs
	2	soft tissue thin	1.25mmx1.25mm	standard		mpr 3d
	3	ankle soft tissue	2.5mmx 2.5mm	standard		pacs
	4	sag bone	2mmx2mm	bone		pacs
	5	coronal bone	2mmx2mm	bone		pacs
	6	sag soft tissue	2mmx2mm	standard		pacs
7	coronal soft tissue	2mmx2mm	standard		pacs	
Scan Start/end location	2cm superior to tibial metadiaphysis					
	2cm inferior calcaneus tuberosity					
	include all of fx and hardware					
DFOV	25 cm					
	decrease appropriately					
3D Technique Used	do 3d spin with recon 2-if fracture seen					
IV contrast volume/type	100ml -isovue 370- if needed for soft tissue infection or mass					
Scan delay	90seconds-Performed as directed by a the supervising radiologist					
Use axial image for sag and coronal reformats						

