

NON CON ABDOMEN/PELVIS 16 Sensation

Indications	For abdomen pain, vomiting, bloating					
Diagnostic Task	Detect diverticulitis, free fluid, appendicitis, obstruction					
Scan mode	Helical					
Position/Landmark	2cm superior to xiphoid/Inspiration					
Topogram	AP 120kV 50mA					
kVp/Reference mass	120kv 160mas/100kv if pt under 140lbs					
Rotation time/pitch	0.5/0.75					
Detector Configuration	16x0.75					
Table Speed/Increment	9					
Dose reduction	CareDose 4D					
Allowed CTDI ranges*	7mGy-50mGy					
XR29 Dose Notification value	50mGy					
Helical Set #1	recon	part	spacing	kernel	window	destination
60 sec delay		body	thickness			recon
	recon	part	spacing	kernel	window	destination
	1	abd/pelvis	2mmx 2mm	31medium smooth	mediastinum	pacs
	2	thin abd/pelvis	1mmx.8mm	31medium smooth	mediastinum	for mpr
	2x2 coronal and sag abd/pelvis reformats from helical set #1, recon 2					
Scan Start/end location	1 cm superior to diaphragm					
	lesser trochanters					
DFOV	40cm					
	decrease appropriately					
IV contrast volume/type	none					
Scan delay	scanned during valsalva if looking for hernia					
	WITH ORAL CONTRAST ONLY					
	Approximate Values for CTDIvol					
	Patient size	weight(kg)	weight(lbs)	CTDIvol(mGy)		
	SMALL	50-70	110-155	10-17		
	AVERAGE	70-90	155-200	15-25		
	LARGE	90-120	200-265	22-35		
NOTE*	<p>*The AAPM recommended NEMA XR29 Dose Notification Value for an adult torso is 50mGy. Dose Notification levels less than the AAPM recommended can be set. The maximum CTDI vol should match the dose notification value. Exams with CTDI vol values less than the minimum allowed range should not be performed unless approved by a radiologist.</p>					