

ROUTINE CHEST/ABDOMEN/PELVIS WO 16 GE

Indications	For abdomen pain, lymphoma, restage ca, weight loss, fatigue				
Diagnostic Task	Detect masses, free fluid, abscess, mets				
Scan mode	Helical				
Position/Landmark	Head first-Supine sternal notch S25-I650				
Topogram	AP 120kV 20mA Lat 120kV 30mA				
kVp/Reference mass	120kv Auto mA (100-440)				
Rotation time/pitch	0.7/1.375:1				
Detector Configuration	16x1.25				
Table Speed/Increment	16.85				
Dose reduction	Noise Index 16.85				
Allowed CTDI ranges*	7mGy-50mGy				
XR29 Dose Notification value	50mGy				
Helical Set 60 sec delay	recon	body part	thickness spacing	algorithm	recon destination
	1	abdomen/pelvis	2.5mmx 2.5mm	standard	paces
	2	lung	1.25mmx1.25mm	lung	paces
	3	sag abdomen	2mmx2mm	standard	paces
	4	coronal abdomen	2mmx2mm	standard	paces
	5	sag chest	2mmx2mm	standard	paces
	6	coronal chest	2mmx2mm	standard	paces
	7	axial MIP lung	10mmx2mm	standard	paces
Scan start/end location	helical set 1 C/A/P-1cm superior to shoulder lesser trochanter				
IV contrast volume/rate	40cm decrease appropriately				
Scan delay	na				
	WITH ORAL MARK AREA OF PAIN WITH BB				
	Approximate Values for CTDIvol				
	Patient size	weight(kg)	weight(lbs)	CTDIvol(mGy)	
	SMALL	50-70	110-155	10-17	
	AVERAGE	70-90	155-200	15-25	
	LARGE	90-120	200-265	22-35	
NOTE*	*The AAPM recommended NEMA XR29 Dose Notification Value for an adult torso is 50mGy. Dose Notification levels less than the AAPM recommended can be set. The maximum CTDI vol should match the dose notification value. Exams with CTDI vol values less than the minimum allowed range should not be performed unless approved by a radiologist.				

