

CT Colonography

64 GE

Indications	Screening				
Diagnostic Task	Detect polyps and colon cancer				
Scan mode	Helical				
Position/Landmark	Head first-Supine Xiphoid S60-I500				
Topogram	AP 120kV 20mA Lat 120kV 40mA				
kVp/Reference mass	120kv Auto mA (300-700)				
Rotation time/pitch	0.5/0.984:1				
Detector Configuration	64x0.625				
Table Speed/Increment	39.37				
Dose reduction	Noise Index 15.86				
Allowed CTDI ranges*	7mGy-50mGy				
XR29 Dose Notification value	50mGy				
Helical Set #1 Supine	recon	body part	thickness spacing	algorithm	recon destination
	1	abdomen/pelvis	2.5mmx 2.5mm	standard	pac
	2	abdomen/pelvis	0.625mmx0.625mm	standard	pac/TR
	3	sag abdomen	2mmx2mm	standard	pac
	4	coronal abdomen	2mmx2mm	standard	pac
Helical Set #2 Prone	recon	body part	thickness spacing	algorithm	recon destination
	1	abdomen/pelvis	2.5mmx 2.5mm	standard	pac
	2	abdomen/pelvis	0.625mmx0.625mm	standard	pac/TR
	3	sag abdomen	2mmx2mm	standard	pac
	4	coronal abdomen	2mmx2mm	standard	pac
Scan Start/end location	1cm superior to diaphragm(include all air) lesser trochanters				
DFOV	40cm decrease appropriately				
IV contrast volume/type	none				
Scan delay	none				
Prep	see prep worksheet				
	see procedure worksheet for CO2 insufflation				

Approximate Values for CTDIvol

Patient size	weight(kg)	weight(lbs)	CTDIvol(mGy)
SMALL	50-70	110-155	10-17
AVERAGE	70-90	155-200	15-25
LARGE	90-120	200-265	22-35

NOTE*

*The AAPM recommended NEMA XR29 Dose Notification Value for an adult torso is 50mGy. Dose Notification levels less than the AAPM recommended can be set. The maximum CTDI vol should match the dose notification value. Exams with CTDI vol values less than the minimum

allowed range should not be performed unless approved by a radiologist.

