

Liver 4 phase+Pelvis 16 Emotion

Indications	New liver lesion with hx of hepatocellular dysfunction or cirrhosis, New HCC, Baseline Cirrohsis, f/u HCC status post TACE or ablation, F/u met disease post ablation					
Diagnostic Task	Detect masses, abscess					
Scan mode	Helical					
Position/Landmark	2cm superior to xiphoid/Inspiration					
Topogram	AP 25mA 130kV					
kVp/Reference mass	130kv 120mas/110kv if pt under 140lbs					
Rotation time/pitch	0.6/0.8					
Detector Configuration	16x1.2					
Table Speed/Increment	15.36					
Dose reduction	CareDose 4D					
Allowed CTDI ranges*	7mGy-50mGy					
XR29 Dose Notification value	50mGy					
Helical Set #1 non con		body	thickness			recon
	recon	part	spacing	kernel	window	destination
	1	abd	2mmx 2mm	31medium smooth	mediastinum	pacs
Helical Set #2 40 sec delay		body	thickness			recon
	recon	part	spacing	kernel	window	destination
	1	abd	2mmx 2mm	31medium smooth	mediastinum	pacs
	2	coronal abdomen	2mmx2mm	31medium smooth	mediastinum	pacs
	3	sag abdomen	2mmx2mm	31medium smooth	mediastinum	pacs
Helical Set #3 70sec		body	thickness			recon
	recon	part	spacing	kernel	window	destination
	1	abd/pelvis	2mmx 2mm	31medium smooth	mediastinum	pacs
	2	coronal abd/pel	2mmx2mm	31medium smooth	mediastinum	pacs
	3	sag abd/pel	2mmx2mm	31medium smooth	mediastinum	pacs
Helical Set #4 5min		body	thickness			recon
	recon	part	spacing	kernel	window	destination
	1	abd	2mmx 2mm	31medium smooth	mediastinum	pacs
	2	coronal abdomen	2mmx2mm	31medium smooth	mediastinum	pacs
	3	sag abdomen	2mmx2mm	31medium smooth	mediastinum	pacs
Scan start	1cm superior to diaphragm					
End location	NC,40sec and 5min delay-iliac crest//// 70sec lesser trochanters					
DFOV	40cm decrease appropriately					
IV contrast volume/rate	75ml < 200lbs, 100ml 200-250lbs, 125ml>250lbs isovue 370 4cc/sec					
Scan delay	Performed as directed by a supervising radiologist					
	noncon/40sec-arterial/ 70sec-venous/5min					
	WITH WATER PREP AND IV CONTRAST					
	Approximate Values for CTDIvol					
	Patient size	weight(kg)	weight(lbs)	CTDIvol(mGy)		
	SMALL	50-70	110-155	10-17		
	AVERAGE	70-90	155-200	15-25		
	LARGE	90-120	200-260	22-30		

NOTE:

*The AAPM recommended NEMA XR29 Dose Notification Value for an adult torso is 50mGy. Dose Notification levels less than the

AAPM recommended can be set. The maximum CTDI vol should match the dose notification value. Exams with CTDI vol values less than the minimum allowed range should not be performed unless approved by a radiologist.

