

# Pancreas 3 phase 16 GE

Indications	For acute pancreatitis, pancreatic mass, pancreatic mass ordered by GI or other subspecialist					
Diagnostic Task	Detect masses, abscess					
Scan mode	Helical					
Position/Landmark	Head first-Supine S25-I500					
Topogram	AP 120kV 10mA Lat 120kV 20mA					
kVp/Reference mass	120kv Smart mA (75-440)					
Rotation time/pitch	0.8/1.375:1					
Detector Configuration	16x1.25					
Table Speed/Increment	27.5					
Dose reduction	Noise Index 15.86					
Allowed CTDI ranges*	7mGy-50mGy					
XR29 Dose Notification value	50mGy					
Helical Set #1 non con	body	thickness				recon
	recon	part	spacing	algorithm	destination	
	1	abdomen	2.5mmx 2.5mm	standard	pacs	
Helical Set #2 40 sec delay	body	thickness				recon
	recon	part	spacing	algorithm	destination	
	1	abdomen	2.5mmx 2.5mm	standard	pacs	
	2	sag abdomen	2mmx2mm	standard	pacs	
	3	coronal abdomen	2mmx2mm	standard	pacs	
Helical Set #3 70 sec delay	body	thickness				recon
	recon	part	spacing	kernel	window	destination
	1	abd	2.5mmx 2.5mm	standard	pacs	
	2	sag abdomen	2mmx2mm	standard	pacs	
	3	coronal abdomen	2mmx2mm	standard	pacs	
Scan start/end location	1cm superior to diaphragm					
for both helical sets	iliac crest					
IV contrast volume/rate	75ml < 200lbs, 100ml 200-250lbs, 125ml>250lbs isovue 370 4cc/sec					
Scan delay	Performed as directed by a supervising radiologist					
	non-con/40sec-arterial/ 70sec-venous					
	Approximate Values for CTDIvol					
	Patient size	weight(kg)	weight(lbs)	CTDIvol(mGy)		
	SMALL	50-70	110-155	10-17		
	AVERAGE	70-90	155-200	15-25		
	LARGE	90-120	200-265	22-35		

NOTE\*

\*The AAPM recommended NEMA XR29 Dose Notification Value for an adult torso is 50mGy. Dose Notification levels less than the

AAPM recommended can be set. The maximum CTDI vol should match the dose notification value. Exams with CTDI vol values less than the minimum allowed range should not be performed unless approved by a radiologist.

