

## Osteomyelitis Pelvis (Other than Sacrum and Coccyx)

**Reviewed By:** Lawrence Tang, MD

**Last Reviewed:** February 2019

**Contact:** (866) 761-4200, Option 1

### General parameters (1.5 T magnets):

For all T1 sequences, please keep TE below 20 (between 10 and 15 if possible); TR 500-600.

For all T2 FS sequences, use equivalent of FSE/TSE. TE of mid to upper 50's is the most ideal for Siemens, 60-65 for GE, and ~ 60 for Toshiba.

It is important to have TE long enough for T2 weighting but not so long that it is signal starved.

For STIR, TI = ~ 135

The purpose of this seemingly complicated approach to osteomyelitis is to streamline the protocol so we can perform the exams on a consistent basis to obtain adequate diagnostic information with a reasonable amount of scanning time.

For all osteomyelitis cases, post-contrast sequences are needed for evaluation of bone viability.

***If intravenous contrast cannot be administered due to severe renal insufficiency or allergy, please refer to routine protocol to scan the patient.***

***Ulcers should be marked before scanning is initiated.***

***Please acquire sequences in the order listed in the protocol.***

If there is difficulty completing the last post-contrast sequence (e.g. pt. motion, pt. pain, scanner shut down etc.), there is no need to repeat the specific sequence.

- whole pelvis
  - ax T1
  - ax STIR
  - cor T1
  - cor STIR
  - ax T1 FS pre contrast
  - ax T1 FS post contrast
  - sag T1 FS post contrast (whole pelvis)