TRA Medical Imaging

TMJ Protocol

Reviewed By: Lawrence Tang, MD **Last Reviewed:** February 2019

Contact: (866) 761-4200, Option 1

General parameters (1.5 T magnets):

For all T1 sequences, please keep TE below 20 (between 10 and 15 if possible); TR 500-600.

For all T2 FS sequences, use equivalent of FSE/TSE. TE of mid to upper 50's is the most ideal for Siemens, 60-65 for GE, and ~ 60 for Toshiba.

It is important to have TE long enough for T2 weighting but not so long that it is signal starved.

For STIR, TI = \sim 135

- CLOSED MOUTH:
 - cor T1
 - sag PD
 - sag T2
- OPEN MOUTH:
 - sag PD
 - sag T2